



Industrie Service

CERTIFICATE



The ATP-Testing Station
of TÜV SÜD Industrie Service GmbH

confirm to the company

Comark Limited
Gunnels Wood Road, Stevenage,
Hertfordshire SG1 2TA, England

because of the positive results of the tested temperature recorder,
type

Comark EVt2 Multi-Use Logger

according to the test-report No. 1059TS dated 28.11.2007

that the temperature recorder fulfils the requirements of the regulation (EC) No 37/2005, the German directive "Zweite Verordnung zur Änderung der Verordnung über tiefgefrorene Lebensmittel" and the standard DIN EN 12830 for the determination of air temperature on means of transport and storage of deep frozen foodstuffs.

The manufacturer is allowed to mark the temperature recorder with the following approval number:

ATP-MUC 1059 TS

Munich, the 28.11.2007

TÜV SÜD Industrie Service GmbH
ATP-Prüfstelle

Birgit Kreß



The recognized
test engineer

Frank Hebenstreit